OLYMPUS AU 480/680/2700/5400

GLUCOSE

HEXOKINASE

Liquid, 2 Reagents

1. General information

Please refer to the package insert for information on reagents, performance characteristics, additionally available products and complete instructions.

Note: Below recommended instrument settings should be used as a guideline in conjunction with respective instrument operation manual.

2. Instrument settings

Test Name	GLUC	Type	Serum	Operation	Yes		
Sample	Volume	1.6	µl	Dilution	0	µl	
Reagents	R1 Volume	120	µl	Dilution	0	µl	
	R2 Volume	30	µl	Dilution	0	µl	
Wavelength	Pri	340		Sec			
Method		END					
Reaction Slope		+					
Measuring point 1	First	0		Last	27		
Measuring point 2	First	0		Last	10		
Linearity			%				
No-Lag-Time							
Pre-dilution Rate							
Min OD				Max OD			
Reagent OD Limit							
	First L	-0.100		First H	2.50		
	Last L	-0.100		Last H	2.50		
Dynamic Range							
	L	0		H	500		
Correlation Factor							
	A	1		B	0		
On-board stability period				#			

Value/Flag	#	Level L	#	Level H	#
Normal Ranges		Age L		Age H	

	Sex	Year	Month	Year	Month
1	#	#	#	#	#
2	#	#	#	#	#
3	#	#	#	#	#
4	#	#	#	#	#
5	#	#	#	#	#
6	#	#	#	#	#
7	#	#	#	#	#

	L	H		
Panic Value	#	#	Unit	mg/dl

Calibration Type	AB	Formula	Y=AX+B	Counts #

	Cal.No	OD	CONC	Factor OD-L	Factor OD-H
Point 1	*		*		
Point 2					
Point 3					
Point 4					
Point 5					
Point 6					
Point 7					

1-Point Cal. Point			
MB Type Factor		Calibration Stability Period	

#) Data entry by the user

*) Enter calibration or standard value and position