DIALAB AUTOLYSER

PROTEIN TOTAL MONO

Biuret
Liquid., Monoreagent

1. General information

Please refer to the package insert for information on reagents, performance characteristics, additionally available products and complete instructions.
Note: Below stated instrument settings should be used as a guideline in conjunction with respective instrument operation manual.

2. Instrument settings

Single reagent application

Code for BAR_ CODE		853
Test Methodology		Biuret Mono
Method		End Point
Kind Of Process		Linear
Filters		578 / 700
Reaction Direction		Increasing
Reagent #1	(µl)	200
Reagent # 2	(µl)	0
Sample Starter		Inactive
Delay Time	(sec)	0
Incubation Time	(sec)	300
Reading Time	(sec)	10
Unit Serum		g/dL
Unit Urine		
Number Of Needle Washes		2
Number Of Cuvette Washes		3
Dynamic Blank		Inactive
Reagent Blank		Every Run
Reagent Limit	(mABS)	200
Curves Acceptance	(%)	100

SERUM PARAMETERS

Name		Protein Total
Sample	(µl)	4
Pre-Dilution		1
Dilution		
Factor		2
Limit Test	(Conc)	15
Low Test Limit	(Conc.)	0.05
Initial ABS	(mABS)	0
Final ABS	(mABS)	0
Max ABS Delta	(mABS)	500
Check Prozona	(mABS)	Inactive
Instrumental Factor		1
Shift		0.000
Re-run Hyperactive		Active
Re-run Pathological		Inactive
RE-run out of curve " Above"		Inactive
Re-run out of curve “Below”		Inactive
Normal Range		
Men		6.60 / 8.80
Women		6.60 / 8.80
Children		5.70 / 8.00